



Study on the transient response of lower limb rehabilitation actuator using the pneumatic cylinder

Dao Minh Duc^{a,*}, Pham Dang Phuoc^a, Tran Xuan Tuy^b, Le Thi Thuy Tram^c

^a Faculty of Engineering Technology, Pham Van Dong University
509 Phan Dinh Phung Street, Quang Ngai City, Viet Nam

^b Faculty of Mechanical, University of Science and Technology
41 Nguyen Luong Bang Street, Lien Chieu, Da Nang City, Viet Nam

^c Department of Electrical and Electronics, Technological Colleges Quang Nam
224 Huynh Thuc Khang, Tam Ky City, Viet Nam

Received 23 January 2018; received in revised form 2 December 2018; accepted 12 December 2018
Published online 30 December 2018

Abstract

A lower limb rehabilitation device was designed using the compressed air cylinder in order to answer the particular request in Viet Nam. This paper is presenting the results of a study of the device response. Dynamic equation of the actuator and equations of the proportional valve have been established. The relationship between the input signal and the output signal of the actuator was derived. Inventor® software was used to design the mechanical structure of the device. Matlab® software was used to calculate the parameters values of the PID controller by simulating the response of the actuator. The results show that the response time of both knee drive and hip drive mechanisms are 8 seconds while the overshoot of both knee drive and hip drive mechanisms are 1%. Moreover, the starting torque of the knee drive mechanism is 17 Nm, and the starting torque of the hip drive mechanism is 35 Nm. The simulation results show that the PID controller gives a fast response time and a low overshoot.

©2018 Research Centre for Electrical Power and Mechatronics - Indonesian Institute of Sciences. This is an open access article under the CC BY-NC-SA license (<https://creativecommons.org/licenses/by-nc-sa/4.0/>).

Keywords: lower limb rehabilitation; hip and knee joint; pneumatic cylinder; PID controller.

I. Introduction

At present, the mechanical engineering industry contributes greatly to the development of society. The products of the mechanical industry are diversified and rich in various fields. In the biomedical field, there are also many products that help the patients and physicians in treating the disease [1]. In the treatment of herniated disc patients with spinal stretch bed, many authors are interested in research, design, manufacture and have achieved certain success [2][3][4].

Some studies have applied robots to support rehabilitation for patients with underlying limb disease, helping patients to shorten treatment and helping doctors to monitor the course of treatment for patients [5]. However, in the treatment of patients with hemiplegia due to a cerebrovascular accident in Viet Nam, at present, there is no automatic device that

supports the treatment of equipment for purely mechanical treatment. Overseas studies have also been available for the treatment of patients and have also been successful [6][7].

Today, compressed air is widely used in industry as well as in social life. The main advantages of compressed air include large capacity, low cost, and ability to be used in harsh environments. These advantages can make the actuator extremely useful in the applications of rehabilitation techniques for stroke or knee pain patients [8][9][10]. The use of compressed air to drive the lower limb rehabilitation device has also been investigated. Previous research had used compressed air to control the actuators and produce very promising results [11][12][13].

The request for equipment to support the treatment of patients with hemiplegia caused by accident has been approved by the researchers. This paper has studied the response of an exercise support structure, which evaluates the feasibility of the structure before practical modeling is expected to contribute to patient treatment

* Corresponding Author. Tel: +84 905 423 314
E-mail address: dmduc@pdu.edu.vn

support. This paper's aim is studying the use of pneumatic cylinders to drive rehabilitation equipment for the hip and knee joint. The advantage of this solution is to build passive exercises for patients with levels 1 to 3 and active impedance exercises for patients with muscle levels of 3 to 5. Combining two active and passive exercises will help the patient shorten the joints rehabilitation time.

II. Materials and methods

A. The lower limb research model

Analysis of the lower limb joints including hip joints, knee joints, and ankle joints shows the total degrees of freedom of the lower limb are 9, as shown in Figure 1 [14]. However, this paper introduces the degrees of hip and knee arthritis that are commonly used in rehabilitation exercises in hospitals and centers nationwide, as illustrated in Figure 2 and Figure 3. Hip and knee joints are reduced to a two-degree system, and the model is shown in Figure 4.

B. Model of lower limb rehabilitation actuator

The functional lower limb rehabilitation actuator model is illustrated in Figure 5. Its structure consists of two stages with two degrees of freedom around the z axis. Stitches are driven by pneumatic cylinders, controlling the cylinders through pneumatic servo valves.




Figure 1. The degrees of freedom of the lower limb



Figure 2. Hip and knee joints in extension state

C. The dynamic equations of the actuator

The dynamic equation of the actuator is expressed by equation (1) and (2).

$$D(q)\ddot{q} + H(q, \dot{q}) + G(q) = \tau \quad (1)$$

$$\tau = \tau_n + \tau_{cc} + \tau_{ms} \quad (2)$$

where τ_n is the torque caused by the hip, and thigh muscles τ_{cc} and τ_{ms} are the friction torque components at the robotic stage.

Because torque components have negligible effects, both components can be ignored and only observing the torque generated by the mechanism. Inertial matrix, centrifugal force and Coriolis matrix, and gravity force matrix are given as follow.

1) Inertial matrix $D(q)$:

$$D_{11} = \frac{1}{3}m_1l_1^2 + m_2\left(l_1^2 + \frac{1}{3}l_2^2\right) + m_2l_1l_2C\theta_2$$

$$D_{12} = D_{21} = \frac{1}{3}m_2l_2^2 + m_2l_1^2 + m_2l_1l_2C\theta_2$$

$$D_{22} = \frac{1}{3}m_2l_2^2$$




Figure 3. Hip and knee joints in the contraction state




Figure 4. Lower limb model




Figure 5. Model of rehabilitation lower limb actuator

2) Centrifugal force and Coriolis matrix:

$$H = \begin{bmatrix} -\frac{1}{2}m_2l_1l_2\dot{\theta}_2^2S\theta_2 - m_2l_1l_2\dot{\theta}_1\dot{\theta}_2S\theta_2 \\ \frac{1}{2}m_2l_1l_2\dot{\theta}_1^2S\theta_2 \end{bmatrix}$$

3) Gravity force matrix:

$$G_1 = \frac{1}{2}m_1gl_1C\theta_1 + \frac{1}{2}m_2gl_1C(\theta_1 + \theta_2) + m_2gl_1C\theta_1$$

$$G_2 = \frac{1}{2}m_2gl_2C(\theta_1 + \theta_2)$$

From the above matrix D, H, G, the dynamical equation (1) becomes:

$$\begin{bmatrix} \tau_1 \\ \tau_2 \end{bmatrix} = \begin{bmatrix} \frac{1}{3}m_1l_1^2 + m_2\left(l_1^2 + \frac{1}{3}l_2^2 + l_1l_2C\theta_2\right) \\ m_2\left(l_1^2 + \frac{1}{3}l_2^2\right) + m_2l_1l_2C\theta_2 \end{bmatrix} \begin{bmatrix} \ddot{\theta}_1 \\ \ddot{\theta}_2 \end{bmatrix} + \begin{bmatrix} -\frac{1}{2}m_2l_1l_2\dot{\theta}_2^2S\theta_2 - m_2l_1l_2\dot{\theta}_1\dot{\theta}_2S\theta_2 \\ \frac{1}{2}m_2l_1l_2\dot{\theta}_1^2S\theta_2 \end{bmatrix} + \begin{bmatrix} \frac{1}{2}m_1gl_1C\theta_1 + \frac{1}{2}m_2gl_1C(\theta_1 + \theta_2) + m_2gl_1C\theta_1 \\ \frac{1}{2}m_2gl_2C(\theta_1 + \theta_2) \end{bmatrix} \quad (3)$$




Figure 6. Model of control at each joint

D. Design of controller for the actuator

Figure 6 shows a model of control mechanism at each joint. At each torque joint, the piston's impact force in the cylinder creates the following torque equation:

$$\tau = \begin{bmatrix} \tau_1 \\ \tau_2 \end{bmatrix} = F.r.\sin\theta \quad (4)$$

where F is the force generated by the pressure in the cylinders affecting the piston area, r is the rotational radius of the link.

Force equation in the cylinder is given by (5).

$$F = \Delta p.A \quad (5)$$

where A is the area of the piston. The valve pressure dependent gas pressure is given by (6).

$$p = K_v.u \quad (6)$$

From equations (3), (4), (5), and estimating point θ , the equation between torque and voltage can be found as follow [15].

$$\tau = \begin{bmatrix} \tau_1 \\ \tau_2 \end{bmatrix} = K_v.r.A.\theta.u \quad (7)$$

PID controller was used to examine and evaluate the response of the structure as shown in Figure 7. Furthermore, equation (1) has been transformed into the equation (8).

$$\ddot{q} = \frac{1}{D(q)}[-H(q,\dot{q}) - G(q)] + \frac{1}{D(q)}\tau \quad (8)$$

The PID controller is described as follow:

$$u = K_p e + K_I \int_0^t e dt + K_D \dot{e} \quad (9)$$

In controlling the actuator, it is necessary to provide the valve opening and closing pressure, from which the pneumatic pressure acts on the piston to produce the torque that controls rotates the rotation at each joint.

The rotation angle of the stages are as follows:

$$\begin{cases} e_1 = \theta_{1d} - \theta_1 \\ e_2 = \theta_{2d} - \theta_2 \end{cases}$$

where θ_{1d} , θ_{2d} are the angles set values of joints 1 and 2. Additionally, denoted control voltage can be found as:




Figure 7. Control chart for the actuator

$$u = \begin{bmatrix} u_1 \\ u_2 \end{bmatrix} \quad (10)$$

Equation (11) can be found by substituting equation (9) into equation (10):

$$\begin{cases} u_1 = K_{P_1} (\theta_{1d} - \theta_1) + K_{I_1} \int_0^t e_1 dt - K_{D_1} \dot{\theta}_1 \\ u_2 = K_{P_2} (\theta_{2d} - \theta_2) + K_{I_2} \int_0^t e_2 dt - K_{D_2} \dot{\theta}_2 \end{cases} \quad (11)$$

The define state variable as follows :

$$\begin{cases} x_1 = \int_0^t e_1 dt \Rightarrow \dot{x}_1 = \theta_{1d} - \theta_1 \\ x_2 = \int_0^t e_2 dt \Rightarrow \dot{x}_2 = \theta_{2d} - \theta_2 \end{cases} \quad (12)$$

Equation (13) can be made by substituting equation (12) into equation (10):

$$\begin{cases} u_1 = K_{P_1} \dot{x}_1 + K_{I_1} x_1 - K_{D_1} \dot{\theta}_1 \\ u_2 = K_{P_2} \dot{x}_2 + K_{I_2} x_2 - K_{D_2} \dot{\theta}_2 \end{cases} \quad (13)$$

Moreover, substituting equation (12) into equation (6) can emerge equation (14):

$$\tau = \begin{bmatrix} K_v \cdot r_1 \cdot A_1 \cdot \theta_1 \cdot (K_{P_1} \dot{x}_1 + K_{I_1} x_1 - K_{D_1} \dot{\theta}_1) \\ K_v \cdot r_2 \cdot A_2 \cdot \theta_2 \cdot (K_{P_2} \dot{x}_2 + K_{I_2} x_2 - K_{D_2} \dot{\theta}_2) \end{bmatrix} \quad (14)$$

Combining equations (8) and (14) will generate this following equation

$$\begin{bmatrix} \ddot{\theta}_1 \\ \ddot{\theta}_2 \end{bmatrix} = \frac{1}{D(q)} [-H(q, \dot{q}) - G(g)] + \frac{1}{D(q)} \begin{bmatrix} K_v \cdot r_1 \cdot A_1 \cdot \theta_1 \cdot (K_{P_1} \dot{x}_1 + K_{I_1} x_1 - K_{D_1} \dot{\theta}_1) \\ K_v \cdot r_2 \cdot A_2 \cdot \theta_2 \cdot (K_{P_2} \dot{x}_2 + K_{I_2} x_2 - K_{D_2} \dot{\theta}_2) \end{bmatrix}$$

III. Results and discussions

A. Device 3D model

After analyzing the dynamics of the lower limb rehabilitation actuator, Inventor® software was used to design mechanical structure for the device. A 3D drawing of the device is shown in Figure 8. The device supports the patient to practice both feet. The patient is able to sit and lie down for training. At the extension and Flexion exercise, the limit angle of the knee is from 0° to 135°. For the hip joint, when the patient is in the sitting position, the limit angle is 0° to 60° in the extension exercise. When the patient is lying down, the limit angle is 0° to 110° at the extension and Flexion exercise.

B. Simulation results

This study was using Simulink® tool in Matlab® to simulate the dynamics of the actuator. The parameter values are selected as follows.

$$A_1 = 0.00524 \text{ (m}^2\text{)};$$

$$A_2 = 0.002826 \text{ (m}^2\text{)};$$

$$l_1 = 0.5 \text{ (m)};$$

$$l_2 = 0.45 \text{ (m)};$$

$$p = 4.105 \text{ (N/m}^2\text{)};$$

$$m_1 = 13 \text{ (kg)}; m_2 = 10 \text{ (kg)};$$

$$K_v = 0,0283 \text{ (N/m}^2 \cdot \text{V)};$$

$$q_0 = \begin{bmatrix} \theta_{01} \\ \theta_{02} \end{bmatrix} = \begin{bmatrix} 0 \\ -\frac{\pi}{2} \end{bmatrix}; \quad q_d = \begin{bmatrix} \theta_{1d} \\ \theta_{2d} \end{bmatrix} = \begin{bmatrix} \frac{\pi}{4} \\ -\frac{\pi}{4} \end{bmatrix}$$

$$K_p = 50;$$

$$K_I = 15;$$

$$K_D = 5;$$




Figure 8. 3D model of the device

Figures 9 to Figure 14 show simulated responses of the actuator. Analysis and discussion of the simulation results are described as follows. In Flexion exercises, the initial values and set point were set using this configuration: the knee joint initial angle was 90° , and the angle set point was 45° , while the hip initial angle was 0° and angle set point was 45° . The following results are obtained:

- 1) The response time of the knee drive mechanism is 8 seconds, and the response time of the hip drive mechanism is 8 seconds.
- 2) Overshoot of the knee drive mechanism is 1%, and overshoot of the hip drive mechanism is 1%.
- 3) The starting torque of the knee drive mechanism is 17 Nm, and the starting torque of the hip drive mechanism is 35 Nm.




Figure 9. Torque response of joint 1




Figure 10. Torque response of joint 2




Figure 11. Angle response of joint 1




Figure 12. Angle response of joint 2




Figure 13. Error angle response of joint 1




Figure 14. Error angle response of joint 2

From the simulation results of the actuator at Flexion exercise for knee and hip, the error is zero and the overshoot is low. However, the response time is long (approximately 6 to 8 seconds), and the starting torque is quite large. These results are due to the

actuator and the human load are quite large. Therefore, the high level of torque is very much required to drive the actuator. From this result, the application of pneumatic cylinder can be used for practical experiments.

IV. Conclusion

In this paper, a dynamic equation for the actuators of a lower limb rehabilitation device was derived, and their responses were simulated using Matlab®. In a Flexion exercise simulation, initial values, and set points were set as: the knee joint initial angle was 90° and its angle setpoint was 45°, while the hip initial angle was 0° and its angle set point was 45°. The results show that the response time for both knee drive and hip drive mechanisms are 8 seconds while the overshoot of both knee drive and hip drive mechanisms are 1%. Moreover, the starting torque of the knee drive mechanism is 17 Nm, and the starting torque of the hip drive mechanism is 35 Nm. The simulation results show that the response is consistent throughout the training of the patient. In the next step, it would be best to examine the empirical model in order to evaluate the results obtained in theoretical simulation.

Acknowledgement

The authors would like to thank the Faculty of the College of Technology, the Pham Van Dong University for cooperation and help.

References

- [1] G.Z Xu, A.G. Song, and H.J. Li, "System design and control technique of robot aided rehabilitation," *Journal of Clinical Rehabilitation Tissue Engineering Research*, vol. 13, page 714-717, 2009.
- [2] Z. Li, D. Guiraud, D. Andreu, A. Gelis, C. Fattal, and M. Hayashibe, "Real-Time closed-loop functional electrical stimulation control of muscle activation with evoked electromyography feedback for spinal cord injured patients," *International Journal of Neural Systems*, vol. 28, no. 06, p. 1750063, Aug. 2018.
- [3] M. O. Ibitoye, N. A. Hamzaid, M. Hayashibe, N. Hasnan, and G. M. Davis, "Restoring prolonged standing via functional electrical stimulation after spinal cord injury: A systematic review of control strategies," *Biomedical Signal Processing and Control*, vol. 49, pp. 34–47, Mar. 2019.
- [4] Y. Chen, J. Hu, W. Wang, L. Peng, L. Peng, and Z.-G. Hou, "An FES-assisted training strategy combined with impedance control for a lower limb rehabilitation robot," in *Proceeding of 2013 IEEE International Conference on Robotics and Biomimetics (ROBIO)*, 2013, pp. 2037–2042.
- [5] P. Coenen, G. Werven, M. Nunen, J. Dieën, K. Gerrits, and T. Janssen, "Robot-assisted walking vs overground walking in stroke patients: An evaluation of muscle activity," *Journal of Rehabilitation Medicine*, vol. 44, no. 4, pp. 331–337, 2012.
- [6] H. Wang, D. Zhang, H. Lu, Y. Feng, P. Xu, R.-V. Mihai, and L. Vladareanu, "Active training research of a lower limb rehabilitation robot based on constrained trajectory," in *Proceeding of 2015 International Conference on Advanced Mechatronic Systems (ICAMEchS)*, 2015, pp. 24–29.
- [7] T. Khoo and S. Senanayake, "Intelligent Musculosoccer Simulator," in *The Impact of Technology on Sport II*, Taylor & Francis, 2007.
- [8] F. Molteni, G. Gasperini, G. Cannaviello, and E. Guanziroli, "Exoskeleton and end-effector robots for upper and lower limbs rehabilitation: Narrative review," *PM&R*, vol. 10, no. 9, pp. S174–S188, Sep. 2018.
- [9] R. R. Torrealba, S. B. Udelman, and E. D. Fonseca-Rojas, "Design of variable impedance actuator for knee joint of a portable human gait rehabilitation exoskeleton," *Mechanism and Machine Theory*, vol. 116, pp. 248–261, Oct. 2017.
- [10] G. Chen, P. Qi, Z. Guo, and H. Yu, "Mechanical design and evaluation of a compact portable knee–ankle–foot robot for gait rehabilitation," *Mechanism and Machine Theory*, vol. 103, pp. 51–64, Sep. 2016.
- [11] J. F. Veneman, R. Kruidhof, E. E. G. Hekman, R. Ekkelenkamp, E. H. F. Van Asseldonk, and H. van der Kooij, "Design and evaluation of the LOPES exoskeleton robot for interactive gait rehabilitation," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 15, no. 3, pp. 379–386, Sep. 2007.
- [12] C. L. Lynch and M. R. Popovic, "A Comparison of closed-loop control algorithms for regulating electrically stimulated knee movements in individuals with spinal cord injury," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 20, no. 4, pp. 539–548, July 2012.
- [13] H. Yu, M. S. Cruz, Gong Chen, Sunan Huang, C. Zhu, E. Chew, Yee Sien Ng, and N. V. Thakor, "Mechanical design of a portable knee-ankle-foot robot," in *2013 Proceeding of IEEE International Conference on Robotics and Automation*, 2013, pp. 2183–2188.
- [14] S. Liu and J. E. Bobrow, "An analysis of a pneumatic servo system and its application to a computer-controlled robot," *Journal of Dynamic Systems, Measurement, and Control*, vol. 110, no. 3, p. 228, 1988.
- [15] H. S. Nam, S. Koh, Y. J. Kim, J. Beom, W. H. Lee, S.-U. Lee, and S. Kim, "Biomechanical reactions of exoskeleton neurerehabilitation robots in spastic elbows and wrists," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 25, no. 11, pp. 2196–2203, Nov. 2017.

This page is intentionally left blank